

Technical

User Guide

IsoClear B™

Product Code: 9302

Search: Astrea Bioseparations

Astrea
Bioseparations

PURITY
by DESIGN

INTRODUCTION

Isoagglutinins are antibodies (IgM and IgG isotypes) which recognise and bind to type A or B blood group antigens. The presence of isoagglutinin antibodies in plasma derived products (particularly plasma for transfusion and IVIG) can give rise to serious side effects, such as haemolysis, depending upon the type of isoagglutinin present (A or B) and the blood type of the recipient. Historically individuals of AB blood type have been used as 'Universal Plasma' donors as their plasma does not contain isoagglutinin antibodies. However, AB is the rarest blood group possessed by a very low percentage of the population. As a result, AB plasma is in great demand and only available in limited quantities. In addition, regulatory authorities are demanding safer IVIG products with low levels of isoagglutinins, to minimise the risk of agglutination and haemolysis in patients receiving IVIG treatments.

Astrea Bioseparations Ltd and its affiliates (Astrea Bioseparations) have developed two affinity chromatography adsorbents (IsoClear A™ & IsoClear B™) comprising of immobilized trisaccharide blood group antigens (A & B). These antigens are chemically synthesized and covalently attached to a biocompatible polymethacrylate support matrix. IsoClear™ adsorbents are highly selective for the removal of isoagglutinin antibodies, enabling cost effective and efficient reduction of isoagglutinin titre for plasma and plasma derived products.

Properties of IsoClear B™:

LIGAND:	Trisaccharide
FUNCTION:	For the capture of anti-B isoagglutinins
ADSORBENT APPEARANCE:	White microspheres
MEAN PARTICLE SIZE (µm):	65
MEAN PORE DIAMETER:	1000 Ångström
MATRIX:	Beaded polymethacrylate resin
RECOMMENDED PACKING CONDITIONS:	Pack at flow rates up to 850 cm/h (pressure up to 2 bar) using 0.1 M NaCl solution
RECOMMENDED OPERATIONAL FLOW RATE:	Up to 500 cm/h
RECOMMENDED OPERATING PRESSURE:	Up to 1 bar
STORAGE*:	2 - 30° C*, 20% ethanol


* **Note:** do not store the resin in buffers of extreme pH or at elevated temperatures.

COLUMN PACKING

1. Before commencing the column pack, consult the relevant manufacturer's instructions for the selected column.
2. Assemble the column and remove air from the dead spaces by flushing the end piece and adaptor with packing solution then close the column outlet.
3. Allow all materials to equilibrate to the temperature at which the chromatography process is to be performed.
4. Slurry the resin and pour the appropriate amount into the column in a single continuous step (**Note:** the material has a compression ratio of ~ 1.15). Pouring the adsorbent down the side of the column helps to prevent air becoming trapped within the adsorbent bed. Ensure that the bottom outlet is not attached to the workstation (into waste vessel) but is closed off.
5. Allow the adsorbent to settle (preferably overnight).
6. Once the adsorbent has settled, ensure the top adaptor is free from air by pumping the packing solution (0.1 M NaCl solution) through it and then place into the top of the column to just above the settled resin. (**Note:** It is recommended to make the buffer on the day of use).
7. Open the bottom outlet and commence the pack under flow.
8. Increase the flow rate up to 850 cm/h (i.e. packing flow rate) or to a maximum pressure of 2 bar (dependant on column dimensions) and pack for 2 CV. Measure and mark the top of the bed. Stop the flow, close the bottom outlet and lower the top adaptor up to 1 cm below the packed bed mark (dependant on bed height), allowing the packing solution out of the top outlet.
9. Open the bottom outlet and recommence packing flow rate for a further 1 CV. There should be no further bed compression (if there is repeat step 8). Slow the flow to operational flow rate and attach the bottom adaptor to the workstation.
10. IsoClear B™ is packed and ready to use.

COLUMN EFFICIENCY TEST

1. Test the column at a flow rate of 100 cm/h.
2. Attach the column to an equilibrated workstation.
3. Commence flow for 1 CV (i.e. to equilibrate and obtain baseline).
4. Inject 2% to 5% CV of a 2% acetone or 2 M NaCl solution.
5. Continue flow until a UV (or conductivity) peak is observed and the trace has returned to baseline (1 to 1.5 CV).
6. End run and determine the asymmetry factor:



7. IsoClear B™ is an affinity adsorbent, therefore an asymmetry factor for an acceptable pack is between 0.8 to 1.6. The recommended plate count for an acceptable pack is ≥ 2000 N/m.

OPERATING INSTRUCTIONS

The following buffers/method is recommended as a starting point for the removal of anti-B isoagglutinins. An initial flow rate of 100 cm/h is recommended. In order to optimise the resin performance (i.e. residence time and column bed height) further optimization may be required.

The preferred option is to use IsoClear B™ with a liquid chromatography system or automated workstation. **Note:** The resin can also be operated manually using peristaltic pumps or even a syringe.

Filter all buffers and feedstock through an appropriate filter, prior to running the column.

1. Equilibrate the column with 5 CV of equilibration solution/buffer (e.g. a near neutral buffer such as PBS - to match the condition of the feedstock). Allow the column buffers and sample to reach the operational temperature.
2. Apply filtered sample onto the column at a flow rate of 100 cm/h. A minimum residence time of 1.5 minutes is required (**Note:** ≥ 3 minutes is recommended).
3. Remove any non-bound material in the column with up to 5 CV of equilibration solution/buffer, or until the UV trace returns to baseline.
4. **If required:** elution of the bound protein may be possible by lowering the pH with up to 5CV of elution buffer (e.g. 50 mM sodium citrate, pH 3.0).
5. It is possible to remove residual adsorbed material by washing the resin with 20% ethanol, 1 M acetic acid as a Clean-in-Place (CIP) step. Alternatively, 6.0 M guanidine HCl can be used.

Note: Avoid alkali buffers/solutions and **DO NOT** use sodium hydroxide (NaOH).

Note: Prolonged contact with the CIP solution may affect the resin performance. Do not store the resin in CIP. Remove the CIP solution with at least 5 bed volumes of equilibration solution/ buffer until the pH and conductivity of the column eluate is equal to that of the buffer at the column inlet.

6. Re-equilibrate the column with 5 CV of equilibration solution/buffer (to remove the CIP solution) and check pH and conductivity before re-use.
7. If required later, store the column in 20% ethanol (2 - 30 °C).

ORDER INFORMATION

Gel Slurry

Code	Description	Pack Size
9302-00025	IsoClear B™	25 mL
9302-00100	IsoClear B™	100 mL
9302-00500	IsoClear B™	500 mL
9302-01000	IsoClear B™	1000 mL

We also offer a range of larger pack sizes for supply of bulk resins into cGMP development and manufacturing scale processes, which are available upon request.

For more information on this or any other supply related matters please do not hesitate to contact us on sales@astrea-bio.com



+44 (0) 1223 433 800 | [astreabioseparations.com](https://www.astreabioseparations.com)

sales@astrea-bio.com | techsupport@astrea-bio.com | quality@astrea-bio.com

Global bases in North America, Canada and Cambridge UK HQ:
Horizon Park, Barton Road, Comberton, Cambridge, CB23 7AJ, UK

Issue Date: 09 May 2022
CCR Number: CCR-1231
Author Name: R Dodd
QA Reviewer Name: R Hawkins

All trademarks, trade names, trade dress, product names and logos appearing in this Technical User Guide are the property of Astrea UK Services Ltd.
Copyright © 2022 Astrea Bioseparations Ltd. All rights are reserved.

Astrea
Bioseparations